

UNIVERSITY COLLEGES OF SCIENCE AND TECHNOLOGY
DEPARTMENT OF CHEMISTRY
UNIVERSITY OF CALCUTTA
92, ACHARYA PRAFULLA CHANDRA ROAD
KOLKATA- 700 009, INDIA



Dr. Susanta Sekhar Adhikari
Professor, Department of Chemistry
University of Calcutta

Phone: 033-2350-8386
FAX: 91 033-2351-9755
E-mail: ssachem@caluniv.ac.in
adhikarisusanta@yahoo.com
Dated: 01.10.2021


Ref no. Chem/SSA/DST/OH/EQP-1/2021-22


Quotation Notice


Sealed quotations for the supply of one electronic analytical balance as per specifications are invited from authorized parties in the following format along with necessary business documents on or before 29.10.2021. The quotation should be typed and addressed to 'Dr. Susanta Sekhar Adhikari, Department of Chemistry, University College of Science & Technology, 92, A.P.C. Road, Kolkata-700009'. The price should be inclusive of relevant taxes/charges, installation, showing break-up. Warranty, on-site service/support should be mentioned clearly. The authority reserves the right to accept/reject any quotations without showing any reason thereof.

Particulars:

Sl. No.	Particulars	Quantity
1	Electronic analytical balance up to 4 decimal precisions equipped with monolithic weighing technology, inbuilt motorized calibration and adjustment with internal weight, 7 segment LCD display. Two tare keys, 4-user selectable filter levels plus optimized display. Weighing capacity: 220 g Readability: 0.0001 g Tare range (subtractive): 220 g Repeatability: ± 0.0001 g Linearity: ± 0.0002 g Pan size (diameter): 90 mm dia. Two tare keys for left and right-handed operation Large glass draft shield with smooth door opening. Dust cover, AC adapter etc.	01 no.


01.10.2021
Dr. Susanta Sekhar Adhikari
Professor
Department of Chemistry
University of Calcutta
92, A.P.C. Road, Kolkata-700 009


01/10/21
Convener, Departmental Purchase Committee
Department of Chemistry
University of Calcutta


01/10/2021
Prof. Chhanda Mukhopadhyay
Head, Department of Chemistry
University of Calcutta